

DEPARTMENT OF HEALTH AND HUMAN SERVICES
CENTERS FOR MEDICARE & MEDICAID SERVICES

PRINTED: 02/11/2021
FORM APPROVED
OMB NO. 0938-0391

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: 49D0988080	(X2) MULTIPLE CONSTRUCTION A. BUILDING _____ B. WING _____	(X3) DATE SURVEY COMPLETED 11/10/2020
NAME OF PROVIDER OR SUPPLIER VIRGINIA LEAGUE FOR PLANNED PARENTHOOD			STREET ADDRESS, CITY, STATE, ZIP CODE 403 YALE DRIVE HAMPTON, VA 23666	
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)	ID PREFIX TAG	PROVIDER'S PLAN OF CORRECTION (EACH CORRECTIVE ACTION SHOULD BE CROSS-REFERENCED TO THE APPROPRIATE DEFICIENCY)	(X5) COMPLETION DATE
D 000	INITIAL COMMENTS	D 000		
D1001	<p>An announced on-site CLIA recertification survey was conducted at Virginia League for Planned Parenthood on November 10, 2020 by the Virginia Department of Health's Office of Licensure and Certification. The survey included an entrance interview on 11/4/2020 and virtual record review conducted on 11/5/2020. The laboratory was surveyed under 42 CFR part 493 CLIA Regulations.</p> <p>Specific deficiencies cited are as follows: CERTIFICATE OF WAIVER TESTS CFR(s): 493.15(e)</p> <p>Laboratories eligible for a certificate of waiver must-- (1) Follow manufacturers' instructions for performing the test; and (2) Meet the requirements in subpart B, Certificate of Waiver, of this part. This STANDARD is not met as evidenced by: Based on a tour, review of manufacturer's package insert, and an interview, the laboratory failed to follow manufacturer's instructions for two (2) of 2 urine pregnancy quality control (QC) vials stored in the laboratory refrigerator on November 10, 2020.</p> <p>Findings include:</p> <p>1. During a laboratory tour on 11/10/20, at approximately 1:00 PM, the inspector noted 2 vials of HCG CON Rapid Response QC (Lot 190916, Level 1 Negative, Level 2 Positive) stored for use in the refrigerator. The vials had manufacturer's printed expiration date of March 16, 2020. The vials did not have an open or</p>	D1001		12/9/20

LABORATORY DIRECTOR'S OR PROVIDER/SUPPLIER REPRESENTATIVE'S SIGNATURE

TITLE

(X6) DATE

12/09/2020

Any deficiency statement ending with an asterisk (*) denotes a deficiency which the institution may be excused from correcting providing it is determined that other safeguards provide sufficient protection to the patients. (See instructions.) Except for nursing homes, the findings stated above are disclosable 90 days following the date of survey whether or not a plan of correction is provided. For nursing homes, the above findings and plans of correction are disclosable 14 days following the date these documents are made available to the facility. If deficiencies are cited, an approved plan of correction is requisite to continued program participation.

DEPARTMENT OF HEALTH AND HUMAN SERVICES
CENTERS FOR MEDICARE & MEDICAID SERVICES

PRINTED: 02/11/2021
FORM APPROVED
OMB NO. 0938-0391

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: 49D0988080	(X2) MULTIPLE CONSTRUCTION A. BUILDING _____ B. WING _____		(X3) DATE SURVEY COMPLETED 11/10/2020
NAME OF PROVIDER OR SUPPLIER VIRGINIA LEAGUE FOR PLANNED PARENTHOOD			STREET ADDRESS, CITY, STATE, ZIP CODE 403 YALE DRIVE HAMPTON, VA 23666		
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)	ID PREFIX TAG	PROVIDER'S PLAN OF CORRECTION (EACH CORRECTIVE ACTION SHOULD BE CROSS-REFERENCED TO THE APPROPRIATE DEFICIENCY)	(X5) COMPLETION DATE	
D1001	Continued From page 1 receive date recorded. The inspector inquired of the testing personnel regarding the protocols for using the observed QC materials. The testing personnel stated: "We use those QC for our urine pregnancy tests". The testing personnel directed the inspector to the facility's supply closet and pointed out one (1) box of corresponding waived HCG CON Rapid Response pregnancy test cassettes stored at room temperature. The test kit contained nine (9) of fifty (50) test cassettes (expiration date January 2021). 2. Review of the BTNX HCG CON Rapid Response QC package insert revealed manufacturer's instructions "Do not use beyond the printed expiration date". 3. In an exit interview with the facility staff at approximately 2:00 PM, the above findings were confirmed.	D1001			
D5215 120M 130M	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2) The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results). This STANDARD is not met as evidenced by: Based on a review of 2019 and 2020 proficiency testing (PT) records and an interview, the	D5215		12/9/20	

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: 49D0988080	(X2) MULTIPLE CONSTRUCTION A. BUILDING _____ B. WING _____		(X3) DATE SURVEY COMPLETED 11/10/2020
NAME OF PROVIDER OR SUPPLIER VIRGINIA LEAGUE FOR PLANNED PARENTHOOD			STREET ADDRESS, CITY, STATE, ZIP CODE 403 YALE DRIVE HAMPTON, VA 23666		
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)	ID PREFIX TAG	PROVIDER'S PLAN OF CORRECTION (EACH CORRECTIVE ACTION SHOULD BE CROSS-REFERENCED TO THE APPROPRIATE DEFICIENCY)	(X5) COMPLETION DATE	
D5215	Continued From page 2 laboratory failed to evaluate non-graded Vaginal Wet Preparation Potassium Hydroxide (KOH) PT results for one (1) of six (6) microscopy events reviewed. Findings include: 1. Review of the laboratory's American Proficiency Institute (API) microscopy PT documentation (2019 Events 1-3, 2020 Events 1-3) revealed no evaluation or verification of accuracy for Vaginal Wet Preparation KOH non-graded response for: 2019 Event 1: challenge sample VKP-01 -not graded due to non consensus. The inspector requested to review the facility's evaluation documentation for the non graded event challenge listed above. No additional documentation was available for review.	D5215			
D5217 120M	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on a review of the Centers for Medicare and Medicaid Services Laboratory Personnel Report Form (CMS 209), proficiency testing (PT) records, and an interview, the laboratory failed to verify the accuracy of Vaginal Wet Preparation	D5217		12/9/20	

STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION		(X1) PROVIDER/SUPPLIER/CLIA IDENTIFICATION NUMBER: 49D0988080	(X2) MULTIPLE CONSTRUCTION A. BUILDING _____ B. WING _____		(X3) DATE SURVEY COMPLETED 11/10/2020
NAME OF PROVIDER OR SUPPLIER VIRGINIA LEAGUE FOR PLANNED PARENTHOOD			STREET ADDRESS, CITY, STATE, ZIP CODE 403 YALE DRIVE HAMPTON, VA 23666		
(X4) ID PREFIX TAG	SUMMARY STATEMENT OF DEFICIENCIES (EACH DEFICIENCY MUST BE PRECEDED BY FULL REGULATORY OR LSC IDENTIFYING INFORMATION)	ID PREFIX TAG	PROVIDER'S PLAN OF CORRECTION (EACH CORRECTIVE ACTION SHOULD BE CROSS-REFERENCED TO THE APPROPRIATE DEFICIENCY)	(X5) COMPLETION DATE	
D5217	<p>Continued From page 3</p> <p>Potassium Hydroxide (KOH) testing twice annually in calendar year 2020.</p> <p>Findings include:</p> <ol style="list-style-type: none"> Review of the laboratory's CMS 209 personnel form, during the entrance interview on 11/5/20, revealed that three (3) testing personnel perform patient Vaginal Wet Preparation (Prep) KOH microscopy examinations. Review of the laboratory's American Proficiency Institute (API) microscopy PT documentation, a total of six (6) events, revealed the lab director utilized API PT to verify Vaginal Wet Prep KOH accuracy twice annually in 2019 and 2020. <p>Review of the PT reports (2019 Events 1-3, 2020 Events 1-3) revealed the laboratory received unsatisfactory Vaginal Wet Prep KOH results for the following two microscopy events:</p> <p>2020 1st Event: 0%; failed for VKP-01; 2020 2nd Event: 0%; failed for VKP-02;</p> <p>resulting in API report statement: "unsuccessful long-term performance".</p> <ol style="list-style-type: none"> In an interview with the chief operating officer on 11/5/20, at approximately 4:00 PM, the above findings were confirmed. 	D5217			