

**GOVERNMENT OF THE DISTRICT OF COLUMBIA
DEPARTMENT OF HEALTH
HEALTH REGULATION ADMINISTRATION
STATEMENT OF DEFICIENCIES AND PLAN OF CORRECTION**

NAME OF FACILITY: Planned Parenthood of Metropolitan Washington	DATE of INSPECTION April 29, 2021
STREET ADDRESS: 1225 4th St NE, Washington, DC 20002	
SUMMARY OF DEFICIENCIES NOTED BY SURVEYING AGENCY	PROVIDER'S PLAN OF CORRECTION WITH TIME TABLE
<p>DC Health conducted the annual Licensure Survey on April 29, 2021. The following findings were based on interviews, record reviews, and observations.</p> <p>2612. Environmental Services and Infection Control.</p> <p>In order to prevent against and provide for the control of infection, each maternity center shall be equipped with the following:</p> <p>This condition is not met as evidenced by:</p> <p>Based on observations during the facility survey, the housekeeping and maintenance services are not adequate to ensure that the facility is maintained in a safe and secure matter.</p> <p>Findings include...</p> <ol style="list-style-type: none"> 1. During the survey, documentation was not available to show that equipment is checked and tested for accuracy by Biomedical Department at least once per year as recommended by the manufacturer. 2. The faucet in the Decontamination Room not secured when examined. 3 The water pressure very low in the Biohazard Room. A small stream of water flowed from the spigot when the surveyor turned the water on. 	<p>Please start typing your responses here.</p> <hr/> <p>Provider's Signature Date</p>